



June 6, 2023

JOHN FRANK  
KI MOBILITY  
5201 WOODWARD DR  
STEVENS POINT, WI 54481

**Document Control Number (DCN): 23069003000001**

Manufacturer Name	Product Name	Model Number	Assigned HCPCS Code(s)
KI MOBILITY	LIBERTY FT	LIBERTY FT	E1161
KI MOBILITY	LIBERTY FT	LIBERTY FT	E1234
KI MOBILITY	LIBERTY - FOLDING	LIBERTY FOLDING BASE	E1161
KI MOBILITY	LIBERTY - FOLDING	LIBERTY FOLDING BASE	E1234
KI MOBILITY	LIBERTY - RIGID	LIBERTY RIGID BASE	E1161
KI MOBILITY	LIBERTY - RIGID	LIBERTY RIGID BASE	E1233

Dear JOHN FRANK,

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles



established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:

E1161 MANUAL ADULT SIZE WHEELCHAIR, INCLUDES TILT IN SPACE

E1233 WHEELCHAIR, PEDIATRIC SIZE, TILT-IN-SPACE, RIGID, ADJUSTABLE, WITHOUT SEATING SYSTEM

E1234 WHEELCHAIR, PEDIATRIC SIZE, TILT-IN-SPACE, FOLDING, ADJUSTABLE, WITHOUT SEATING SYSTEM

In order for a wheelchair to be determined to have a seating system, the base unit must already include features that directly support or contain the body of the occupant. Please see ISO 7176-26 for definition of a seating support system.

If you disagree with this decision, you may request a reconsideration within 45 days of the letter's date and provide evidence to substantiate a reconsideration of PDAC's original coding determination. To request a reconsideration, complete the Reconsideration Request form located on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). If your request for a reconsideration is made after the 45-day time frame, it will require a new application and documentation to support the request.

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the Product Classification List (PCL) on the Durable Medical Equipment Coding System (DMECS). Further information for requesting updates to the PCL can be found on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Palmetto GBA; nor does it imply or guarantee claim reimbursement or coverage.

If you have questions, please contact the PDAC HCPCS Helpline at (877) 735-1326 during the hours of 9:30 a.m. to 5:00 p.m. ET, Monday through Friday. You may also visit our [website](#) to chat with one of our representatives or select the Contact Us button at the top of the page for email, FAX or postal mail information.

Sincerely,

Pricing, Data Analysis, and Coding (PDAC)  
Palmetto GBA, LLC  
[www.dmepdac.com](http://www.dmepdac.com)